Protocol for Real-Time RT-PCR assay with Lyoph P&P ZIKV Duo (adapted to TaqMan platform)

For RUO (Research Use Only)

Vial content

Lyoph P&P ZIKV Duo contains lyophilized primers and probe (P&P) for detection of Zika virus (ZIKV). It must be used in combination with a generic RT-qPCR kit.

Design

RT-qPCR systems included in Lyoph P&P ZIKV Duo:

- Assay optimization for molecular detection of Zika Virus. Corman *et al.*, Bull World Health Organ, 2016. <u>DOI</u>: 10.2471/BLT.16.175950
- In house assay

Caution

Vials containing primers and probe mix (Lyoph P&P) must be stored at -20°C in the dark after reception. Stable for at least 6 months under the described conditions.

Instructions

This protocol is adapted to the following kit and RT-qPCR platform:

- SuperScript™ III Platinum™ One-Step qRT-PCR Kit (ThermoFisher)
- CFX (BIORAD) RT-qPCR thermal cycler

Use of other kits or RT-qPCR platforms may require adaptation of the protocol.

I. Rehydration of Lyoph-P&P

- Write the date on the vial before opening
- Resuspend Lyoph P&P as follows (these steps are critical to ensure adequate homogenization):
 - ✓ Add nuclease-free water or RE buffer as described in Table 1
 - ✓ Homogenize by pipetting 10 to 20 times the volume up and down in the glass vial
 - ✓ Incubate rehydrated P&P at room temperature for 10 min
 - ✓ Perform a second series of 10-times pipetting.

Table 1. Lyoph-P&P regeneration; the number of tests/vial is indicated on the Lyoph P&P vial label.

Packaging (number of						
tests/vial)	8+1	16+2	24+2	32+3	48+3	96+5
H₂O or RE buffer (μL)	63	126	182	245	357	707

II. <u>Preparation of the PCR reaction mix (for SuperScript™ III Platinum™ One-Step qRT-PCR Kit - ThermoFisher)</u>

Table 2. Preparation of the PCR mix. *Rehydration performed as indicated in Table 1.

	Volume/PCR test (μL)
2X Reaction mix	12.5
Rehydrated Lyoph P&P*	7
SSIII/Taq enzyme mix	0.5
Total (1 PCR test)	20
Template RNA (H ₂ O for Ctrl-)	5
Final volume	25

III. Cycling program and RT-qPCR (for SuperScript™ III Platinum™ One-Step qRT-PCR Kit - ThermoFisher)

50°C	15min	
95°C	2min	
95°C	15sec	45 cycles
60°C	30sec (plate read)	45 Cycles

Read: FAM-QSY

Rehydrated Lyoph-P&P can be stored at 4 °C in the dark for up to 7 days.